



SAFETY AND BUILDINGS DIVISION
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

July 31, 2012

EMANUELSON-PODAS, INC
JOHN NORDSTROM
10401 BREU ROAD EAST
MINNETONKA MN 55343

SURGERY CENTER SERVICES OF AMERICA
SCOTT BLAIR
325 LAKE FRONT DR
ORLANDO FL 98843

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: SURGERY CENTER SERVICES OF AMERICA
Product Name: PRE/POST VAC STEAM AUTOCLAVE
Model Number(s): AMSCO #3013
Product File No: 20120265

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of July 2017.

This approval is contingent upon compliance with the following stipulation(s):

- The indirect waste pipe from this product must discharge by means of an air-gap, an air-gap fitting that conforms to ASME A112.1.3 or an air-gap fitting approved by the Department.
- A reduced pressure principle backflow preventer (ASSE 1013 or CAN/CSA B64.4) or an assembly or method suitable for high hazard must be installed on the water supply pipes serving this product.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Jerry Thompson
Plumbing Product Reviewer
Phone: 608-266-6742
Fax: 608-267-9566
E-mail: jerry.thompson@wi.gov